



THE AMERICAN ASSOCIATION FOR  
LABORATORY ACCREDITATION

## ACCREDITED LABORATORY

A2LA has accredited

**ALABAMA SCALE & INSTRUMENT – MARSHALL, INC.**  
**d.b.a. ASI Calibration Labs - Texas**  
**Marshall, TX**

for technical competence in the field of **Calibration**

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2005 *General Requirements for the Competence of Testing and Calibration Laboratories*. This laboratory also meets the requirements of ANSI/NCSL Z540-1-1994 and any additional program requirements in the field of calibration. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (*refer to joint ISO-ILAC-IAF Communiqué dated 18 June 2005*).

Presented this 24<sup>th</sup> day of 2008.

A handwritten signature in cursive script, reading "Peter Abney", positioned above a horizontal line.

President  
For the Accreditation Council  
Certificate Number 1876.02  
Valid to February 28, 2010  
REVISED June 25, 2008



For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005  
& ANSI/NCSL Z540-1-1994

ALABAMA SCALE & INSTRUMENT – MARSHALL, INC.  
d.b.a. ASI Calibration Labs - Texas  
3314 Monigold Road  
Marshall, TX 75672  
Lawson Hill Phone: 251 660 9999

CALIBRATION

Valid To: February 28, 2010

Certificate Number: 1876.02

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations<sup>1</sup>:

I. Dimensional

| Parameter/Equipment                                      | Range                                     | Best Uncertainty <sup>2, 5</sup> ( $\pm$ )                             | Comments                              |
|--|---|--|---------------------------------------|
| Calipers, Verniers <sup>3</sup>                          | Up to 36 in                               | 290 $\mu$ in   | Gage blocks and rod standards         |
| Height Gages <sup>3</sup>                                | Up to 36 in                               | 0.16 % of reading  | Gage blocks and rod standards         |
| Micrometers <sup>3</sup> –<br>Outside<br>Inside<br>Depth | Up to 36 in<br>Up to 36 in<br>Up to 36 in | (110 + 0.94L) $\mu$ in<br>(110 + 16L) $\mu$ in<br>(110 + 16L) $\mu$ in | Supermic, gage blocks and comparator. |
| Dial and Test Indicators <sup>3</sup>                    | Up to 2 in                                | 250 $\mu$ in   | Gage blocks                           |
| Pin Gages  | Up to 6 in                                | (4.7 + 9.1L) $\mu$ in  | Supermic., gage blocks and comparator |

| Parameter/Equipment   | Range                                    | Best Uncertainty <sup>2,5</sup> ( $\pm$ )  | Comments  |
|---|--|--|---|
| Plain Plug Gages  | Up to 6 in                               | $(4.7 + 9.1L) \mu\text{in}$  | Supermic, gage blocks, and comparator                                       |
| Verification of Length <sup>3</sup> – Rules and Tapes   | Up to 360 in<br>Up to 1200 in            | $(0.012 + 0.00008L)$ in<br>$(0.0085 + 0.016L)$ in                                    | Gage blocks, linear standards   |
| Gage Blocks   | (0.3125 to 1) in<br>2 in<br>3 in<br>4 in | 3.6 $\mu\text{in}$<br>4.3 $\mu\text{in}$<br>5.1 $\mu\text{in}$<br>6.1 $\mu\text{in}$ | Gage block comparator   |
| Long Gage Blocks  | (5 to 20) in                             | $(6.4 + 1.2L) \mu\text{in}$  | Comparison of gage blocks using Brown & Sharpe digital probe with amplifier |
| Thread Measuring Wires  | (4 to 120) TPI                           | 13 $\mu\text{in}$  | Gage blocks and Supermic  |
| Length Standards –<br>Setting Rods  | Up to 40 in                              | $(6.4 + 1.2L) \mu\text{in}$  | Gage blocks, Brown & Sharpe digital probe with amplifier                    |
| Thread Plugs –<br>Standard 60° Screw,<br>Metric Screw<br><br>Major Diameter<br>Pitch Diameter           | Up to 6 in<br>Up to 6 in                 | $(7.7 + 4L) \mu\text{in}$<br>160 $\mu\text{in}$                                      | Supermic with thread measuring wires  |
| Thread Plugs –<br>Pipe Inch (NPT,<br>NPSM, NPSL, ANPT,<br>NPTF)<br><br>Major Diameter<br>Pitch Diameter | Up to 6 in<br>Up to 6 in                 | $(10 + 3.8L) \mu\text{in}$<br>170 $\mu\text{in}$                                     | Supermic with thread measuring wires  |

| Parameter/Equipment   | Range   | Best Uncertainty <sup>2,5</sup> ( $\pm$ ) | Comments                           |
|---|---|---|------------------------------------|
| Thread Rings –<br><br>Standard 60° Screw,<br>Metric Screw Pipe Inch<br>(NPT, NPSM, NPSL,<br>ANPT, NPTF) GO /<br>NO-GO | Up to 3 in  | 300 $\mu$ in                              | Master threaded set<br>plug        |
| Ring Gages  | (0.001 to 10) in                                    | (20 + 5L) $\mu$ in                        | Supermicrometer and<br>gage blocks |
| Surface Plates –<br>Repeatability Only  | (18 in $\times$ 24 in) to<br>(72 in $\times$ 96 in) | 50 $\mu$ in                               | Repeat-o-meter                     |

## II. Mechanical

| Parameter/Equipment   | Range                                      | Best Uncertainty <sup>2,5</sup> ( $\pm$ ) | Comments   |
|---|--|---|--|
| Balances <sup>3,4</sup>   | 1 mg to 13.2 kg                            | 0.5 <i>LSD</i>                            | Verification with<br>Class 1 weights   |
| Scales and Belt Scales <sup>4</sup>                             | (0.005 to 200 000) lb                      | 0.5 <i>LSD</i>                            | Verification with<br>Class 6 and Class F<br>weights per NIST<br>HB44               |
| Pressure Gages,<br>Transducers and<br>Transmitters <sup>3</sup> | (-15 to 300) psi<br><br>(50 to 10 000) psi | 0.084 psi<br><br>0.2 % of reading         | ASME B40, Druck<br>610<br><br>Deadweight tester                                    |
| Verification of Force <sup>3</sup> –<br>Tension and Compression | (0 to 10 000) lbf                          | 0.1 % of reading                          | Verification by NIST<br>HB44 and ASTM E4<br>with Class F weights<br>and load cells |

| Parameter/Equipment   | Range   | Best Uncertainty <sup>2</sup> (±)   | Comments                          |
|---|---|---|-----------------------------------|
| Torque <sup>3</sup>   | (4 to 50) in·lb<br>(30 to 400) in·lb<br>(80 to 100) in·lb<br><br>(20 to 250) ft·lb<br>(100 to 1000) ft·lb     | 0.15 in·lb<br>1.2 in·lb<br>2.9 in·lb<br><br>0.73 ft·lb<br>3.1 ft·lb                       | Torque transducers and indicators |
| Indirect Verification of Rockwell Hardness Testers              | HRB<br><br>Low (43.31)<br>Middle (65.13)<br>High (82.50)<br><br>Low (43.31)<br>Middle (65.13)<br>High (82.50) | 1 HRB<br>0.64 HRB<br>0.46 HRB<br><br>Repeatability:<br>0.24 HRB<br>0.32 HRB<br>0.28 HRB   | ASTM E18                          |
| Indirect Verification of Rockwell Hardness Testers <sup>3</sup> | HRC<br><br>Low (27.53)<br>Middle (45.60)<br>High (62.48)<br><br>Low (27.53)<br>Middle (45.60)<br>High (62.48) | 0.38 HRC<br>0.33 HRC<br>0.31 HRC<br><br>Repeatability:<br>0.21 HRC<br>0.28 HRC<br>0.2 HRC | ASTM E18                          |

### III. Thermodynamics

| Parameter/Equipment               | Range  | Best Uncertainty <sup>2,3</sup> (±)    | Comments                       |
|-----------------------------------|--|--|--------------------------------|
| Temperature – Measure             | (-200 to 850) °C                               | 0.46 °C + 9.25 x 10 <sup>-5</sup> / °C | Omega CL-27 with Pt 100 Ω RTD  |
| Temperature – Measuring Equipment | (Ambient + 5 to 100) °C<br><br>(100 to 300) °C | 0.42 °C<br><br>0.78 °C                 | Hart Scientific 9100 dry block |
| Ice Point                         | 0 °C   | 0.2 °C                                 | ASTM E563-02                   |

| Parameter/Equipment  | Range  | Best Uncertainty <sup>2,3</sup> (±)  | Comments                           |
|--|--|--|------------------------------------|
| RTD – Simulation<br><br>Pt 385, 100 Ω<br>Pt 385, 1000 Ω  | <br><br>(-200 to 800) °C<br>(-200 to 630) °C   | <br><br>0.49 °C<br>0.39 °C   | <br><br>Fluke 725                  |
| Thermocouple<br>Simulation –<br><br>Type K<br><br>Type J<br><br>Type T<br><br>Type E<br><br>Type N<br><br>Type S | <br><br>(-200 to 0) °C<br>(0 to 1370) °C<br><br>(-200 to 0) °C<br>(0 to 1200) °C<br><br>(-200 to 0) °C<br>(0 to 400) °C<br><br>(-200 to 0) °C<br>(0 to 950) °C<br><br>(-20 to 0) °C<br>(0 to 500) °C<br>(500 to 1750) °C<br><br>(-20 to 0) °C<br>(0 to 500) °C<br>(500 to 1750) °C | <br><br>1.6 °C<br>1.2 °C<br><br>1.4 °C<br>1.1 °C<br><br>1.6 °C<br>1.2 °C<br><br>1.3 °C<br>1.1 °C<br><br>3.0 °C<br>2.2 °C<br>1.8 °C<br><br>3.0 °C<br>2.2 °C<br>1.8 °C | <br><br>Fluke 725                  |
| Relative Humidity –<br>Measure   | (0 to 90) % RH<br><br>(90 to 100) % RH   | 3 % RH<br>4.4 % RH   | Vaisala HMI41 w<br>HMP42           |
| Ovens, Furnaces, and<br>Environmental<br>Chamber Uniformity  | (Ambient +20 to 500) °C  | 0.5 °C   | ASTM E145-01                       |
| Infrared Temperature<br>Measuring Equipment  | Ambient to 100°C<br>(100 to 500) °C  | 1.3 °C<br>2.0 °C   | Hart Scientific 9132<br>black body |

#### IV. Time and Frequency

| Parameter/Equipment             | Range              | Best Uncertainty <sup>2,3</sup> ( $\pm$ ) | Comments   |
|---------------------------------|--------------------|---|--|
| Digital / Mechanical Tachometer | (40 to 99 000) RPM | 0.1 % of reading + LSD                    | Direct reflective pickup tachometer<br><br>LSD = least significant digit |
| Timers                          | 15 s to 24 hr      | 0.53 s                                    | Reference stopwatch  |
| Stopwatches                     | 15 s to 24 hr      | 0.53 s                                    | NIST synchronized computer clock   |

<sup>1</sup> This laboratory offers commercial calibration services and field calibration services (where noted).

<sup>2</sup> “Best Uncertainty” is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards of nearly ideal measuring equipment. Best uncertainties represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of  $k = 2$ . The best uncertainty of a specific calibration performed by the laboratory may be greater than the best uncertainty due to the behavior of the customer’s device and to influences from the circumstances of the specific calibration.

<sup>3</sup> Field calibration service is available for this calibration. The uncertainties achievable on a customer’s site can normally be expected to be larger than the Best Measurement Capabilities (BMC) that the accredited laboratory has been assigned as Best Uncertainty on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the uncertainty introduced by the item being calibrated, (e.g. resolution) must also be considered and this, on its own, could result in the calibration uncertainty being larger than the BMC.”

<sup>4</sup> The uncertainty of scale verification is highly dependent on local conditions such as the resolution of the scale. Any statement of best uncertainty would therefore be misleading. The class of the best weights used by the laboratory is shown in the Comments column.

<sup>5</sup> In the statement of best uncertainty,  $L$  is the numerical value of the nominal length of the device measured in inches, TPI is threads per inch, and LSD is least significant digit.